STERILIZATION RECORDS

Harbour Dental Care

PERIOD
13 Oct 2026 – 26 Oct 2026

STANDARD FOLLOWED
CDC Guidelines

GENERATED
27 Oct 2026, 09:12

Records in Clavio are locked once saved, so entries cannot be edited or deleted after the fact. Later clarifications are appended as dated notes.
Protocol: spore test every week · chemical indicator on every load · printout check optional · records kept 3 years.

56
Cycles logged

1
Failed indicators · 1 resolved, 0 open

4 / 4
Spore tests passed

4
Consumable lots traced

1 Sterilization cycles

56

DATE / TIME	AUTOCLAVE	CYCLE	PROGRAM / LOAD	INDICATOR / PRINTOUT	STRIP LOT	OP
13 Oct 08:00	Autoclave A	#1180	Universal 134 °C Handpieces Tray 1, 2 handpiece pouches	Pass Printout reached 134.1 °C · 4 min	CI-2311	EW
13 Oct 09:23	Autoclave B	#412	Universal 134 °C Surgical kit Extraction kit	Pass Printout reached 134.2 °C · 4 min	CI-2311	JR
13 Oct 11:46	Autoclave A	#1181	Quick 134 °C Mixed Hygiene set, 3 pouches	Pass Printout reached 134.3 °C · 4 min	CI-2311	DK
13 Oct 12:09	Autoclave B	#413	Prion 134 °C Wrapped Tray 4, scalers	Pass Printout reached 121.3 °C · 15 min	CI-2311	EW
13 Oct 14:32	Autoclave A	#1182	Universal 134 °C Impression trays	Pass	CI-2311	JR
13 Oct 15:55	Autoclave B	#414	Universal 134 °C Handpieces Endo set	Pass Printout reached 134.3 °C · 4 min	CI-2311	DK
14 Oct 08:00	Autoclave A	#1183	Quick 134 °C Surgical kit	Pass Printout reached 134.1 °C · 4 min	CI-2311	EW

DATE / TIME	AUTOCLAVE	CYCLE	PROGRAM / LOAD	INDICATOR / PRINTOUT	STRIP LOT	OP
14 Oct 09:23	Autoclave B	#415	Universal 134 °C Mixed Tray 2, 4 pouches	Pass Printout reached 121.3 °C · 15 min	CI-2311	JR
14 Oct 11:46	Autoclave A	#1184	Universal 134 °C Wrapped Tray 1, 2 handpiece pouches	Pass Printout reached 134.3 °C · 4 min	CI-2311	DK
14 Oct 12:09	Autoclave B	#416	Prion 134 °C Impression trays Extraction kit	Pass	CI-2311	EW
14 Oct 14:32	Autoclave A	#1185	Quick 134 °C Handpieces Hygiene set, 3 pouches	Pass Printout reached 134.2 °C · 4 min	CI-2311	JR
14 Oct 15:55	Autoclave B	#417	Universal 134 °C Surgical kit Tray 4, scalers	Pass Printout reached 121.3 °C · 15 min	CI-2311	DK
15 Oct 08:00	Autoclave A	#1186	Universal 134 °C Mixed	Pass Printout reached 134.1 °C · 4 min	CI-2311	EW
15 Oct 09:23	Autoclave B	#418	Universal 134 °C Wrapped Endo set	Pass Printout reached 134.2 °C · 4 min	CI-2311	JR
15 Oct 11:46	Autoclave A	#1187	Quick 134 °C Impression trays	Pass	CI-2311	DK
15 Oct 12:09	Autoclave B	#419	Prion 134 °C Handpieces Tray 2, 4 pouches	Pass Printout reached 121.3 °C · 15 min	CI-2311	EW
15 Oct 14:32	Autoclave A	#1188	Universal 134 °C Surgical kit Tray 1, 2 handpiece pouches	Pass Printout reached 134.2 °C · 4 min	CI-2311	JR
15 Oct 15:55	Autoclave B	#420	Universal 134 °C Mixed Extraction kit	Pass Printout reached 134.3 °C · 4 min	CI-2311	DK
16 Oct 08:00	Autoclave A	#1189	Quick 134 °C Wrapped Hygiene set, 3 pouches	Pass Printout reached 134.1 °C · 4 min	CI-2311	EW
16 Oct 09:23	Autoclave B	#421	Universal 134 °C Impression trays Tray 4, scalers	Pass	CI-2311	JR

DATE / TIME	AUTOCLAVE	CYCLE	PROGRAM / LOAD	INDICATOR / PRINTOUT	STRIP LOT	OP
16 Oct 11:46	Autoclave A	#1190	Universal 134 °C Handpieces	Pass Printout reached 134.3 °C · 4 min	CI-2311	DK
16 Oct 12:09	Autoclave B	#422	Prion 134 °C Surgical kit Endo set	Pass Printout reached 134.1 °C · 4 min	CI-2311	EW
16 Oct 14:32	Autoclave A	#1191	Quick 134 °C Mixed	Pass Printout reached 134.2 °C · 4 min	CI-2311	JR
16 Oct 15:55	Autoclave B	#423	Universal 134 °C Wrapped Tray 2, 4 pouches	Pass Printout reached 121.3 °C · 15 min	CI-2311	DK
17 Oct 08:00	Autoclave A	#1192	Universal 134 °C Impression trays Tray 1, 2 handpiece pouches	Pass	CI-2311	EW
17 Oct 10:23	Autoclave B	#424	Universal 134 °C Handpieces Extraction kit	Pass Printout reached 134.2 °C · 4 min	CI-2311	JR
17 Oct 12:46	Autoclave A	#1193	Quick 134 °C Surgical kit Hygiene set, 3 pouches	Pass Printout reached 134.3 °C · 4 min	CI-2311	DK
17 Oct 14:09	Autoclave B	#425	Prion 134 °C Mixed Tray 4, scalers	Pass Printout reached 121.3 °C · 15 min	CI-2311	EW
20 Oct 08:00	Autoclave A	#1194	Universal 134 °C Wrapped	Pass Printout reached 134.2 °C · 4 min	CI-2402	EW
20 Oct 09:23	Autoclave B	#426	Universal 134 °C Impression trays Endo set	Pass	CI-2402	JR
20 Oct 11:46	Autoclave A	#1195	Quick 134 °C Handpieces	Pass Printout reached 134.1 °C · 4 min	CI-2402	DK
20 Oct 12:09	Autoclave B	#427	Universal 134 °C Surgical kit Tray 2, 4 pouches	Pass Printout reached 121.3 °C · 15 min	CI-2402	EW
20 Oct 14:32	Autoclave A	#1196	Universal 134 °C Mixed Tray 1, 2 handpiece pouches	Pass Printout reached 134.3 °C · 4 min	CI-2402	JR

DATE / TIME	AUTOCLAVE	CYCLE	PROGRAM / LOAD	INDICATOR / PRINTOUT	STRIP LOT	OP
20 Oct 15:55	Autoclave B	#428	Prion 134 °C Wrapped Extraction kit	Pass Printout reached 134.1 °C · 4 min	CI-2402	DK
21 Oct 08:00	Autoclave A	#1197	Quick 134 °C Impression trays Hygiene set, 3 pouches	Pass	CI-2402	EW
21 Oct 09:23	Autoclave B	#429	Universal 134 °C Handpieces Tray 4, scalers	Pass Printout reached 121.3 °C · 15 min	CI-2402	JR
21 Oct 11:46	Autoclave A	#1198	Universal 134 °C Surgical kit	Fail · resolved Printout NOT reached 128.4 °C · 2 min	CI-2402	DK
21 Oct 12:09	Autoclave B	#430	Universal 134 °C Mixed Endo set	Pass Printout reached 134.2 °C · 4 min	CI-2402	EW
21 Oct 14:32	Autoclave A	#1199	Quick 134 °C Wrapped	Pass Printout reached 134.3 °C · 4 min	CI-2402	JR
21 Oct 15:55	Autoclave B	#431	Prion 134 °C Impression trays Tray 2, 4 pouches	Pass	CI-2402	DK
22 Oct 08:00	Autoclave A	#1200	Universal 134 °C Handpieces Tray 1, 2 handpiece pouches	Pass Printout reached 134.2 °C · 4 min	CI-2402	EW
22 Oct 09:23	Autoclave B	#432	Universal 134 °C Surgical kit Extraction kit	Pass Printout reached 134.3 °C · 4 min	CI-2402	JR
22 Oct 11:46	Autoclave A	#1201	Quick 134 °C Mixed Hygiene set, 3 pouches	Pass Printout reached 134.1 °C · 4 min	CI-2402	DK
22 Oct 12:09	Autoclave B	#433	Universal 134 °C Wrapped Tray 4, scalers	Pass Printout reached 121.3 °C · 15 min	CI-2402	EW
22 Oct 14:32	Autoclave A	#1202	Universal 134 °C Impression trays	Pass	CI-2402	JR
22 Oct 15:55	Autoclave B	#434	Prion 134 °C Handpieces Endo set	Pass Printout reached 134.1 °C · 4 min	CI-2402	DK

DATE / TIME	AUTOCLAVE	CYCLE	PROGRAM / LOAD	INDICATOR / PRINTOUT	STRIP LOT	OP
23 Oct 08:00	Autoclave A	#1203	Quick 134 °C Surgical kit	Pass Printout reached 134.2 °C · 4 min	CI-2402	EW
23 Oct 09:23	Autoclave B	#435	Universal 134 °C Mixed Tray 2, 4 pouches	Pass Printout reached 121.3 °C · 15 min	CI-2402	JR
23 Oct 11:46	Autoclave A	#1204	Universal 134 °C Wrapped Tray 1, 2 handpiece pouches	Pass Printout reached 134.1 °C · 4 min	CI-2402	DK
23 Oct 12:09	Autoclave B	#436	Universal 134 °C Impression trays Extraction kit	Pass	CI-2402	EW
23 Oct 14:32	Autoclave A	#1205	Quick 134 °C Handpieces Hygiene set, 3 pouches	Pass Printout reached 134.3 °C · 4 min	CI-2402	JR
23 Oct 15:55	Autoclave B	#437	Prion 134 °C Surgical kit Tray 4, scalers	Pass Printout reached 121.3 °C · 15 min	CI-2402	DK
24 Oct 08:00	Autoclave A	#1206	Universal 134 °C Mixed	Pass Printout reached 134.2 °C · 4 min	CI-2402	EW
24 Oct 10:23	Autoclave B	#438	Universal 134 °C Wrapped Endo set	Pass Printout reached 134.3 °C · 4 min	CI-2402	JR
24 Oct 12:46	Autoclave A	#1207	Quick 134 °C Impression trays	Pass	CI-2402	DK
24 Oct 14:09	Autoclave B	#439	Universal 134 °C Handpieces Tray 2, 4 pouches	Pass Printout reached 121.3 °C · 15 min	CI-2402	EW

2 Weekly spore tests

4

STARTED	AUTOCLAVE	KIT LOT	RESULT	READ	BY	FAILURE PROTOCOL
13 Oct 07:40	Autoclave A	BI-7731	Pass	**14 Oct** 08:05 Photo on file	EW	—
13 Oct 07:55	Autoclave B	BI-7731	Pass	**14 Oct** 08:07 Photo on file	EW	—
20 Oct 07:45	Autoclave A	BI-7731	Pass	**21 Oct** 08:10 Photo on file	JR	—
20 Oct 07:58	Autoclave B	BI-7731	Pass	**21 Oct** 08:12 Photo on file	JR	—

3 Corrective actions

1

DATE	CYCLE	ACTION TAKEN	LEAD NOTIFIED	STATUS
21 Oct 11:52	Autoclave A #1198	Load quarantined & re-run Re-run as cycle #1203. Strip from the same lot passed on the re-run.	Yes	Resolved 21 Oct 2026, 12:05 Cause: Pouch folded over the strip; chamber loaded too tightly

4 Consumable lots

4

ITEM	LOT	EXPIRES	ADDED	STATUS	USED
Chemical indicator strips	CI-2311	30 Nov 2026	2 Aug 2026	Replaced 20 Oct	28
Chemical indicator strips	CI-2402	30 Sep 2027	20 Oct 2026	In use	28
Spore test kits	BI-7731	31 Mar 2027	1 Sep 2026	In use	4
Sterilization pouches	SP-11820	30 Jun 2028	1 Oct 2026	In use	0

5 Autoclaves & servicing

2

AUTOCLAVE	SERIAL	CLASS	LAST SERVICED	EVERY	CALIBRATION DUE	STATUS
Autoclave A Tuttnauer EZ11Plus	EZ11-4471-20	B	6 Oct 2026	12 mo	6 Oct 2027	In service
Autoclave B Midmark M11	M11-889-0931	N	14 Apr 2026	12 mo	14 Apr 2027	In service

6 Service & repair log

1

DATE	AUTOCLAVE	REASON	TECHNICIAN	WORK DONE	OUTCOME
6 Oct 2026	Autoclave A	Routine service	J. Doe, Coastal Sterilizer Service Report photo on file	Annual service. Door gasket replaced, chamber temperature verified at 134 °C, safety valve tested. Parts: Door gasket	Back in service Next 6 Oct 2027

7 Team

3

INITIALS	NAME	ROLE
DK	Dr Alex Kim	Dentist
EW	Emma Walsh	Practice admin
JR	Jordan Reyes	Dental assistant

Verification hash (SHA-256 of the records above) `28164bbaaa3b52fe389e20beb988c50d48265ff0812195e889f4a561519bf8a6`
Generated by Clavio for Harbour Dental Care on 27 Oct 2026, 09:12. Re-exporting the same period from unchanged records produces the same hash.
Photos marked "on file" are stored with the record and can be shown from the app on request.